

TOTAL IRON TEST

(Sensitivity 0,5 to 10 mg/l(ppm) Fe)

1. Rinse the test tube with the water to be tested and fill the 5ml. Mark.
2. Add 5 drops Total Iron Indicator, stopper the test tube and shake gently.
3. After 3-5 minutes, compare the reddish color to color reference chart to determine concentration of total iron in mg/l (ppm)
4. Color comparison must be completed within 20 minutes.